

RNAscope® 2.5 HD Detection Kit (BROWN) Quick Guide

For FFPE Tissues

Introduction

This quick guide is intended for advanced users who are familiar with the procedures in the *Formalin-Fixed Paraffin-Embedded (FFPE) Sample Preparation and Pretreatment User Manual, Part 1* (Document No. 322452-USM) and *RNAscope® 2.5 HD Detection Reagent BROWN User Manual, Part 2* (Document No. 322310-USM). Refer to the user manual for safety guidelines. For every chemical, read the Safety Data Sheet (SDS) and follow handling instructions. Wear appropriate protective eyewear, clothing, and gloves. For the latest services and support information, go to: www.acdbio.com/support.

Part 1: Prepare and Pretreat Samples

Workflow Steps	
PREPARE FFPE SECTIONS	<ol style="list-style-type: none"> 1. Immediately place dissected tissue sample in fresh 10% NBF for 16–32 HRS at ROOM TEMPERATURE (RT). 2. Dehydrate, embed in paraffin, and cut sample into 5 +/- 1 µm sections. Mount sections on Superfrost® Plus slides. <hr/> OPTIONAL STOPPING POINT (1). Use sectioned tissue within 3 months. Store sections with dessicants at RT.
PREPARE SLIDES ~1.5 HOURS Bake Slides ↓ Deparaffinize FFPE Sections	Bake Slides <ol style="list-style-type: none"> 1. Bake slides in a dry oven for 1 HR at 60°C. <hr/> OPTIONAL STOPPING POINT (2). Use sectioned tissue within 1 week. Store sections with dessicants at RT. <hr/> Deparaffinize FFPE Sections <ol style="list-style-type: none"> 1. In a fume hood: <ul style="list-style-type: none"> • Fill two Tissue-Tek® Clearing Agent dishes with ~200 mL fresh xylene. • Fill two Tissue-Tek® Staining dishes with ~200 mL fresh 100% EtOH. 2. Place slides in a Tissue-Tek® Slide Rack in xylene 2 x 5 MIN. 3. Incubate slides in 100% EtOH 2 x 1 MIN. 4. Remove slides from rack. Air dry slides for 5 MIN at RT. <hr/> OPTIONAL STOPPING POINT (3). Air dry overnight at RT (must use within 24 hrs) or proceed directly to the next step.
PRETREAT SAMPLES ~1–2 HOURS Prepare Oven and Reagents ↓ Apply H2O2 (Pretreat 1) ↓ Apply Target Retrieval (Pretreat 2) ↓ Create Barrier ↓	Prepare Oven and Reagents (30 MIN at 40°C) <ol style="list-style-type: none"> 1. Set HybEZ™ oven to 40°C and warm HybEZ™ Humidity Control Tray containing wet Humidifying Paper for 30 MIN before use. Keep tray warm during assay. 2. Prepare 700 mL fresh 1X Target Retrieval in a beaker. Cover with foil, bring to a mild boil, and maintain. Do not boil more than 30 MIN before use. Apply RNAscope® Hydrogen Peroxide (10 MIN at RT) <ol style="list-style-type: none"> 1. Add ~5–8 drops of Hydrogen Peroxide to each section for 10 MIN at RT. 2. Place slides into a Tissue-Tek® Slide Rack submerged in distilled water. 3. Wash slides in the distilled water by moving the rack up and down 3–5 times and repeat with fresh distilled water. Apply RNAscope® Target Retrieval

<p>Apply Protease Plus (Pretreat 3)</p>	<ol style="list-style-type: none"> 1. With a pair of forceps <i>very slowly</i> submerge the slide rack into boiling 1X Target Retrieval solution. Refer to Appendix A of the <i>Formalin-Fixed Paraffin-Embedded (FFPE) Sample Preparation and Pretreatment User Manual, Part 1</i> (Catalog No. 320511) for specific pretreatment time, depending on your tissue type. 2. <i>Immediately</i> transfer hot slide rack to a staining dish containing distilled water. 3. Wash slides in the distilled water by moving the rack up and down 3–5 times and repeat with fresh distilled water. 4. Wash slides in fresh 100% EtOH by moving the rack up and down 3–5 times, and air dry. <p>Create Barrier</p> <ol style="list-style-type: none"> 1. Draw 2–4 times around tissue using the Immedge™ hydrophobic barrier pen. Dry completely ~2 MIN or OVERNIGHT at RT. <p>Apply Protease Plus</p> <ol style="list-style-type: none"> 1. Place slides in the HybEZ™ Slide Rack, and add ~5 drops of Protease Plus to each section. 2. Place the HybEZ™ Slide Rack in the prewarmed HybEZ™ Humidity Control Tray. Seal tray and insert back into the HybEZ™ Oven. Incubate at 40°C for 30 MIN. <p>Note: If needed, prepare RNAscope® 2.5 assay materials during this step.</p> <ol style="list-style-type: none"> 3. Wash slides in the distilled water by moving the rack up and down 3–5 times and repeat with fresh distilled water.
--	--

Part 2: RNAscope® 2.5 Assay

Workflow Steps	
<p>PREPARE THE MATERIALS ~10–30 MIN</p>	<ol style="list-style-type: none"> 1. Prepare 3 L of 1X Wash Buffer by adding 2.94 L distilled water and 1 bottle (60 mL) of 50X Wash Buffer to a large carboy. Mix well. Warm 50X Wash Buffer up to 40°C for 10–20 MIN before making 1X Wash Buffer. 2. Prepare 50% Hematoxylin and 0.02% Ammonia water. 3. Prepare dehydrating reagents: 200 mL xylene in a clearing agent dish, 2 x 200 mL 100% ETOH and 200 mL 70% ETOH in staining dishes. 4. Equilibrate reagents and equipment: <ul style="list-style-type: none"> • Remove Amp 1–6 from refrigerator. • Warm probes for 10 MIN at 40°C and cool to RT.
<p>RUN THE ASSAY ~5 HOURS</p> <p>Hybridize Probe ↓ Hybridize Amp 1 ↓ Hybridize Amp 2 ↓ Hybridize Amp 3 ↓ Hybridize Amp 4 ↓ Hybridize Amp 5 ↓</p>	<p>Hybridize Probe (2 HRS at 40°C)</p> <ol style="list-style-type: none"> 1. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and add ~4 drops probe to each section. 2. Insert sealed tray containing HybEZ™ Slide Rack back into the HybEZ™ Oven for 2 HRS at 40°C. Remove slide rack. 3. Wash slides in 1X Wash Buffer for 2 MIN at RT. Repeat with fresh 1X Wash Buffer. <p>Hybridize Amp 1 (30 MIN at 40°C)</p> <ol style="list-style-type: none"> 1. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and add ~4 drops Amp 1 to each section. 2. Insert sealed tray containing HybEZ™ Slide Rack into the HybEZ™ Oven for 30 MIN at 40°C. Remove slide rack. 3. Wash slides in 1X Wash Buffer for 2 MIN at RT. Repeat with fresh 1X Wash Buffer. <p>Hybridize Amp 2 (15 MIN at 40°C)</p> <ol style="list-style-type: none"> 1. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and add ~4 drops Amp 2 to each section.

<p style="text-align: center;">Hybridize Amp 6 ↓ Detect the Signal ↓ Counterstain the Slides ↓ Mount the Slides</p>	<ol style="list-style-type: none"> 2. Insert sealed tray containing HybEZ™ Slide Rack into the HybEZ™ Oven for 15 MIN at 40°C. Remove slide rack. 3. Wash slides in 1X Wash Buffer for 2 MIN at RT. Repeat with fresh buffer. <p style="color: #A52A2A; margin-top: 0;">Hybridize Amp 3 (30 MIN at 40°C)</p> <ol style="list-style-type: none"> 1. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and add ~4 drops Amp 3 to each section. 2. Insert sealed tray containing HybEZ™ Slide Rack into the HybEZ™ Oven for 30 MIN at 40°C. Remove slide rack. 3. Wash slides in 1X Wash Buffer for 2 MIN at RT. Repeat with fresh buffer. <p style="color: #A52A2A; margin-top: 0;">Hybridize Amp 4 (15 MIN at 40°C)</p> <ol style="list-style-type: none"> 1. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and add ~4 drops Amp 4 to each section. 2. Insert sealed tray containing HybEZ™ Slide Rack into the HybEZ™ Oven for 15 MIN at 40°C. Remove slide rack, but do <i>not</i> place tray back into the oven. 3. Wash slides in 1X Wash Buffer for 2 MIN at RT. Repeat with fresh buffer. <p style="color: #A52A2A; margin-top: 0;">Hybridize Amp 5 (30 MIN at RT)</p> <ol style="list-style-type: none"> 1. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and add ~4 drops Amp 5 to each section. 2. Incubate sealed tray containing HybEZ™ Slide Rack for 30 MIN at RT. 3. Wash slides in 1X Wash Buffer for 2 MIN at RT. Repeat with fresh buffer. <p style="color: #A52A2A; margin-top: 0;">Hybridize Amp 6 (15 MIN at RT)</p> <ol style="list-style-type: none"> 1. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and add ~4 drops Amp 6 to each section. 2. Incubate sealed tray containing HybEZ™ Slide Rack for 15 MIN at RT. 3. Wash slides in 1X Wash Buffer for 2 MIN at RT. Repeat with fresh buffer. <p style="color: #A52A2A; margin-top: 0;">Detect the Signal (10 MIN at RT)</p> <ol style="list-style-type: none"> 1. Mix equal volumes of BROWN-A and BROWN-B. 2. Remove excess liquid from slides, place in the HybEZ™ Slide Rack, and pipette ~120 µL of DAB onto each tissue section. 3. Incubate sealed tray containing HybEZ™ Slide Rack for 10 MIN at RT. 4. Remove DAB from slides and wash 3–5 times in distilled water. <p style="color: #A52A2A; margin-top: 0;">Counterstain the Slides (2 MIN at RT)</p> <ol style="list-style-type: none"> 1. Place slides in 50% Hematoxylin I for 2 MIN at RT. Wash 3–5 times in distilled water and repeat with fresh distilled water. 2. Wash slides 10 SEC in 0.02% Ammonia water, and then wash 3–5 times in distilled water. <p style="color: #A52A2A; margin-top: 0;">Mount the Slides</p> <ol style="list-style-type: none"> 1. Incubate slides in 70% EtOH for 2 MIN with occasional agitation. 2. Incubate slides in 95% EtOH for 2 MIN with occasional agitation. Repeat with fresh EtOH. 3. Incubate slides in xylene for 5 MIN with occasional agitation. 4. Add 1–2 drops of Cytoseal and place coverslip over section and air dry.
<p>EVALUATE THE RESULTS</p>	<p>Examine tissue sections under a standard bright field microscope at 20–40X magnification</p>

Troubleshooting

For troubleshooting information, please contact technical support at support@acdbio.com.

For Research Use Only. Not for Diagnostic Use.

NOTICE TO PURCHASER: PLEASE REFER TO THE RNASCOPE® 2.5 ASSAY- USER MANUAL FOR LIMITED USE LABEL LICENSE OR DISCLAIMER INFORMATION. Advanced Cell Diagnostics, Inc. reserves the right to change its products and services at any time to incorporate technological developments. This manual is subject to change without notice. Although this manual has been prepared with every precaution to ensure accuracy, Advanced Cell Diagnostics, Inc. assumes no liability for any errors, omissions, or for any damages resulting from the use of this information.

© 2016 Advanced Cell Diagnostics. All rights reserved. RNAscope® and HybEZ™ are trademarks of Advanced Cell Diagnostics, Inc. All other trademarks belong to their respective owners.

Headquarters

7707 Gateway Blvd Suite 200, Newark, CA 94545 Phone 1-510-576-8800 Toll Free 1-877-576-3636

For support, email support@acdbio.com

www.acdbio.com

