

# RNAscope® VS Target Probe

**Description**

The RNAscope® VS Target Probe is designed to work with the RNAscope® VS Reagent Kit for semi and fully automated *in situ* detection of RNA in formalin-fixed paraffin-embedded (FFPE) tissue specimens on slides using the Discovery™ XT and ULTRA Systems. Each kit contains sufficient reagents in a convenient, Ready-To-Use (RTU) format to perform 30 slides total. The results of the assay can be viewed using a standard bright field microscope.

**Contents and Storage**

The components of the RNAscope® VS Target Probes (Cat. No. Various) and quantities supplied are listed in the following table. The contents have a shelf life until the date of expiry indicated on the reagent label. It is important to follow the product storage conditions as specified on the label so that the reagents maintain their reliability.

Catalog No.	Component	Quantity	Storage
Various	RNAscope® VS Target Probe [species]	11 mL x 1 bottle	2-8 °C

**IMPORTANT!** Before running the assay please download and review online RNAscope® VS Reagent Kit-BROWN or RED User Manual at <http://www.acdbio.com/support/technical-doc>.

Note: Some tissue types may require modified pretreatment conditions. Please refer to online tissue pretreatment guidelines: <http://www.acdbio.com/go>, Contact [support@acdbio.com](mailto:support@acdbio.com) for suggestions or questions.

**Precautions**

Handle biological specimens and other materials according to local laboratory guidelines and regulations. Avoid cross-contamination of samples and reagents to ensure quality results. Use equipment and reagents as specified by the manufacturer.

For safety and biohazard guidelines, refer to the "Safety" section in the RNAscope® VS Reagent Kit-BROWN and/or RED User Manual. For every chemical, read the Material Safety Data Sheets (MSDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves. MSDS are available online at

<http://www.acdbio.com/support/technical-doc/category/msds>.

**Contact Information**

**Address:** 3960 Point Eden Way, Hayward, CA 94545 USA  
**Phone:** 1-877-576-3636 option 1 **Fax:** 1-510-576-8798  
**Web:** [www.acdbio.com](http://www.acdbio.com) **Email:** [support@acdbio.com](mailto:support@acdbio.com)

**Copyright**

©2015, Advanced Cell Diagnostics, Inc. All rights reserved.

**Trademarks**

RNAscope® is a trademark of Advanced Cell Diagnostics, Inc. Leica BOND RX is a registered trademark of Leica Biosystems. All other trademarks belong to their respective owners.

Advanced Cell Diagnostics, Inc. reserves the right to change its products and services at any time to incorporate technological developments. This insert is subject to change without notice. Although this insert has been prepared with every precaution to ensure accuracy, Advanced Cell Diagnostics, Inc. assumes no liability for any errors, omissions, or for any damages resulting from the use of this information.