



RNAscope® 2.5 LS Reagent Kit – RED Product Insert

Description

The RNAscope® 2.5 LS Reagent Kit- RED (Cat. No. 322150) is designed to work with the RNAscope® 2.5 LS Target Probes (Cat. No. 300038) for fully automated *in situ* detection of RNA in formalin-fixed paraffin-embedded (FFPE) tissue specimens on slides using the Leica BOND RX instrument. Each kit contains sufficient reagents in a convenient, Ready-To-Use (RTU) format to perform 60 slides total. The results of the assay can be viewed using a standard bright field microscope.

Contents and Storage

The components of the RNAscope® 2.5 LS Reagent Kit—RED and quantities supplied are listed in the following table. The reagents have a shelf life from the date of manufacture to the expiration date as printed on the kit label. It is important to follow the product storage conditions as specified on the label so that the reagents maintain their reliability.

Catalog No.	Component	Quantity	Storage
322150	RNAscope® 2.5 LS Hydrogen Peroxide	21 mL x 1 bottle	2 – 8°C
	RNAscope® 2.5 LS Protease III	21 mL x 1 bottle	2 – 8°C
	RNAscope® 2.5 LS Rinse	29 mL x 2 bottles	2 – 8°C
	RNAscope® 2.5 LS AMP 1	21 mL x 1 bottle	2 – 8°C
	RNAscope® 2.5 LS AMP 2	21 mL x 1 bottle	2 – 8°C
	RNAscope® 2.5 LS AMP 3	21 mL x 1 bottle	2 – 8°C
	RNAscope® 2.5 LS AMP 4	21 mL x 1 bottle	2 – 8°C
	RNAscope® 2.5 LS AMP 5 - RED	21 mL x 1 bottle	2 – 8°C
	RNAscope® 2.5 LS AMP 6 - RED	21 mL x 1 bottle	2 – 8°C

IMPORTANT! The RNAscope® 2.5 LS Assay requires the RNAscope® 2.5 LS Probes and the RNAscope® 2.5 LS Reagents, available from Advanced Cell Diagnostics.

IMPORTANT! Before running the assay please download and review online RNAscope® 2.5 LS Reagent Kit- RED User Manual (Cat. No. 322150-USM) at <http://www.acdbio.com/technical-support/downloads>.

Note: Some tissue types may require modified pretreatment conditions. Please refer to online tissue pretreatment guidelines: <http://www.acdbio.com/go>. Contact support@acdbio.com for suggestions or questions.

Precautions

Handle biological specimens and other materials according to local laboratory guidelines and regulations. Avoid cross-contamination of samples and reagents to ensure quality results. Use equipment and reagents as specified by the manufacturer. For safety and biohazard guidelines, refer to the "Safety" section in the RNAscope® 2.5 LS Reagent Kit-RED User Manual. For every reagent, read the Safety Data Sheets (SDSs) and follow the handling instructions available at <http://www.acdbio.com/support/technical-doc/category/sds>. Wear appropriate protective eyewear, clothing, and gloves.

Contact Information

Address: 3960 Point Eden Way, Hayward, CA 94545 USA
Phone: 1-877-576-3636 option 1 **Fax:** 1-510-576-8801
Web: www.acdbio.com **Email:** support@acdbio.com

Copyright ©2015, Advanced Cell Diagnostics, Inc. All rights reserved.

Trademarks
RNAscope® is a trademark of Advanced Cell Diagnostics, Inc. Leica BOND RX is a registered trademark of Leica Biosystems. All other trademarks belong to their respective owners. Advanced Cell Diagnostics, Inc. reserves the right to change its products and services at any time to incorporate technological developments. This insert is subject to change without notice. Although this insert has been prepared with every precaution to ensure accuracy, Advanced Cell Diagnostics, Inc. assumes no liability for any errors, omissions, or for any damages resulting from the use of this information.